V.5/Oct/2011



Rat Leptin ELISA Kit Instructions

For the quantitative determination of leptin in rat serum, plasma and fluid

> Catalog #: 90040 96 Assays

For Research Use Only. Not for use in diagnostic procedures.

1536 Brook Drive, Suite A Downers Grove, IL 60515, USA Tel: (630) 889-9003 Fax: (630) 889-9021 E-mail: sales@crystalchem.com URL: http://www.crystalchem.com

TABLE OF CONTENTS

А.	Intended Use	1
В.	Introduction	1
С.	Principles of the Assay	2
D.	Kit Storage	2
E.	Assay Materials	
	E.1. Materials supplied	3
	E.2. Materials required but not provided	3
F.	Reagent Precautions	4
G.	Maximizing Kit Performance	4
H.	Preparation of Rat Plasma and Serum	4
/.	Assay (0.2 – 12.8 ng/mL)	
	I.1. Preparation of reagents	5
	I.2. Preparation of working rat leptin standards	6
	I.3. Assay procedure	7
	I.4. Determining the leptin concentration	8
J.	Appendix	
	J.1. Performance characteristics	9
	J.2. Rat leptin recovery test	9
	J.3. Summary of reagent preparation	10
	J.4. Summary of Rat Leptin ELISA kit assay	11
Wa	arranty	12

A. Intended Use

The Rat Leptin ELISA kit is for the quantitative determination of leptin in rat serum, plasma, and fluid. Please read the complete kit insert before performing this assay. The kit is for *RESEARCH USE ONLY*. It is not intended for use in clinical or diagnostic procedures or for internal or external use in humans or animals.

B. Introduction

Obesity is a significant contributing factor in various adult diseases such as diabetes, cardiac disease, etc. This fact, combined with the increasing prevalence of obesity in the human population, has resulted in increased research on the underlying impact and cause of obesity.

In 1994, leptin, *obese* gene product, was identified from the investigation of ob/ob mouse. Leptin is a protein of about 16 kDa, which is expressed in adipose tissue, and promotes weight loss by suppressing appetite and stimulating metabolism. As a result, the accurate measurement of leptin in experimental animals is becoming increasingly important as obesity research intensifies.

The kit is a simple, precise, and sensitive ELISA sandwich assay for rat leptin.

Sensitivity range of the assay: 0.2 to 12.8 ng/mL using 5µl sample.

C. Principles of the Assay

1. First reaction

Rat leptin in the sample is simultaneously bound to the rabbit anti-leptin antibody coated on the microplate well and the anti-leptin IgG of the guinea pig anti-serum added to each well.

2. Washing

Unbound material is removed by washing.

3. Second reaction

Horse radish peroxidase (POD)-conjugated anti-guinea pig IgG antibody is then bound to the guinea pig anti-leptin IgG of the complex immobilized to the microplate well.

4. Washing

Excess POD-conjugated antibody is removed by washing.

5. Enzyme reaction

The bound POD conjugated antibody in the microplate well is detected by the addition of the 3, 3', 5, 5'-tetramethylbenzidine (TMB) substrate solution.

6. Measurement of absorbance

7. Evaluation of results

The leptin concentration is determined via interpolation using the standard curve generated by plotting absorbance versus the corresponding concentration of rat leptin standard.

D. Kit Storage

- 1. Upon receipt of the Rat Leptin ELISA kit, store it at 2-8°C and avoid light exposure (do not freeze the kit or hold it at temperatures above 25°C).
- 2. The kit should not be used after the expiration date.

E. Assay Materials E.1. Materials supplied

Mark	Description	Amount				
	Antibody-coated Microplate					
Α	(One pack contains 6×8 well modules,	2 packs				
	i.e., 48 wells / pack)					
В	Pot Lontin Standard Lyanhilizad	2.56 ng/vial				
D	Rat Leptin Standard, Lyophilized	(for 100 µL)				
С	Sample Diluent	1 bottle				
C	Sample Diluent	(50 mL)				
D	Guinea Pig Anti-Leptin Serum	1 bottle				
		(6 mL)				
Е	Anti-Guinea Pig IgG Enzyme Conjugate Stock	1 bottle				
L	Solution	(8.4 mL)				
F	Enzyme Conjugate Diluent	1 bottle				
-	Enzyme Conjugate Dildent	(3.6 mL)				
G	Enzyme Substrate (TMB) Solution	1 bottle				
0		(13 mL)				
н	Enzyme Reaction Stop Solution	1 bottle				
	(1 N Sulfuric acid)	(13 mL)				
	Wash Buffer Stock Solution	1 bottle				
	(20X concentrate)	(50 mL)				
	Frame for affixing the microplate well module	1 piece				
	Plastic microplate cover	1 piece				

E.2. Materials required but not provided

Micropipettes and disposable tips Volumetric flasks Distilled or deionized water Polypropylene microtubes Test tube racks Vortex mixer Aspirator for washing procedure Microplate reader (capable of measuring A₄₅₀ and A₆₃₀ values)

F. Reagent Precautions

- 1. Avoid direct contact with the Enzyme Substrate Solution (marked "G") and the Enzyme Reaction Stop Solution (marked "H"). In case of contact, immediately flush eyes or skin with plenty of water and get medical advice.
- 2. Do not allow the Enzyme Substrate Solution (marked "G") to contact any metal.
- 3. Only appropriately-trained personnel should use the kit. Laboratory personnel should wear suitable protective clothing. All chemicals should be considered potentially hazardous.

G. Maximizing Kit Performance

- 1. Given the small sample volumes required (5 μ L), pipetting should be done as carefully as possible. A high quality 10 μ L or better precision pipette should be used for such volumes. Drops of liquid adhering to the outside of the pipette tips should be removed by wiping to ensure the highest degree of accuracy.
- 2. In order to prevent the microplate wells from drying, samples and reagents should be dispensed quickly into the wells. In no case should 10 minutes be exceeded per plate per pipetting step.
- 3. The wash procedure should be done thoroughly in order to minimize background readings.
- 4. Each standard and sample should be assayed in duplicate.
- 5. The same sequence of pipetting and other operations should be maintained in all procedures.
- 6. Do not mix reagents that have different lot numbers.

H. Preparation of Rat Plasma and Serum

- **Plasma:** Collect blood into a tube containing an anticoagulant such as heparin (final concentration: 1 unit/mL), EDTA (final concentration: 0.1%), or sodium citrate (final concentration: 0.76%), and centrifuge for 20 min at 2,000 x *g*.
- Serum: Collect blood, allow to clot, and centrifuge for 20 min at 2,000 x g.
- **Note:** Be sure to avoid hemolysis during preparation. Do not use turbid serum or plasma samples. Turbid serum or plasma should be centrifuged to produce a clear solution. <u>Samples</u> which need to be diluted must be diluted using the Sample <u>Diluent (marked "C").</u>

I. Assay (0.2 – 12.8 ng/mL)

I.1. Preparation of reagents

Prior to use, all reagents should be brought to room temperature (18-25°C), and should be stored at 2-8°C immediately after use. Before use, mix the reagents thoroughly by gentle agitation or swirling.

1. Antibody-coated microplate

Remove the "Antibody-coated Microplate" (marked "A") from the sealed foil pouch after the pouch has been equilibrated to room temperature.

Note: The microplate must be used the same day as the pouch is opened.

2. Rat leptin stock solution

Reconstitute the "Rat Leptin Standard, Lyophilized" (marked "B") by careful addition of 100 μ L of sample diluent to the vial. Invert the vial gently until the contents are completely dissolved. This stock solution contains 25.6 ng/mL of rat leptin. The reconstituted rat leptin stock solution is stable for one week at 2-8°C.

3. Sample diluent

The "Sample Diluent" (marked "C") is provided as a ready-touse preparation. Once the bottle is opened, the sample diluent is stable for one week at 2-8°C.

4. Guinea pig anti-leptin serum

The "Guinea Pig Anti-Leptin Serum" (marked "D") is provided as a ready-to-use preparation. Once the bottle is opened, the guinea pig anti-leptin serum is stable for one week at 2-8°C.

5. Anti-guinea pig IgG enzyme conjugate

Mix the bottle of "Anti-Guinea Pig IgG Enzyme Conjugate Stock Solution" (marked "E") with the bottle of "Enzyme Conjugate Diluent" (marked "F"). Mix completely to ensure a homogeneous and <u>clear</u> solution. Avoid foaming during mixing. The anti-guinea pig IgG enzyme conjugate is stable for one week at 2-8°C.

Note: The anti-guinea pig IgG enzyme conjugate is not needed till the second day of the assay.

6. Enzyme substrate solution

The "Ezyme Substrate Solution" (marked "G") is provided as a ready-to-use preparation. Once the bottle is opened, the enzyme substrate solution is stable for one week at 2-8°C.

Note: Avoid exposure of the enzyme substrate solution to light.

7. Enzyme reaction stop solution (1 N sulfuric acid)

The "Enzyme Reaction Stop Solution" (marked "H") is provided as a ready-to-use preparation.

8. Wash buffer

The "Wash Buffer Stock Solution" (marked "I") should be brought to 1 L with distilled or deionized water in a volumetric flask. Mix the solution well before use. The wash buffer is stable for one week at 2-8°C.

I.2. Preparation of working rat leptin standards

- Pipette 50 μL of sample diluent (marked "C") and 50 μL of rat leptin stock solution (25.6 ng/mL) into a polypropylene microtube labeled 12.8 ng/mL, and mix thoroughly.
- 2. Dispense 50 µL of sample diluent into six polypropylene microtubes labeled 0.2, 0.4, 0.8, 1.6, 3.2 and 6.4 ng/mL, respectively.
- 3. Dispense 50 µL of the 12.8 ng/mL standard into the 6.4 ng/mL microtube, and mix thoroughly.
- 4. Dispense 50 μL of the 6.4 ng/mL standard into the 3.2 ng/mL microtube, and mix thoroughly.
- 5. Repeat this dilution scheme using the remaining microtubes.
- 6. Dispense 50 μL of sample diluent into one polypropylene microtube labeled 0 ng/mL.
 - **Note**: The working leptin standards should be prepared shortly before use and discarded after use. Prepare working leptin standards using polypropylene microtubes because polypropylene exhibits minimal adsorption of leptin

TABLE 2 Preparation of working rat leptin standards								
	Rat leptin concentration (ng/mL)							
	12.8	6.4	3.2	1.6	0.8	0.4	0.2	0
RLSS*	50µL							
SD**	50µL	50µL	50µL	50µL	50µL	50µL	50µL	50µL
		50µL	50µL	50µL	50µL	_50μL	_ 50μL	
	/	• /	/	/	/	• 7		
Total	100µL	100µL	100µL	100µL	100µL	100µL	100µL	50µL

TABLE 2 Preparation of working rat leptin standards

RLSS*: Rat Leptin Stock Solution (25.6 ng/mL) SD**: Sample Diluent

I.3. Assay Procedure

First reaction:

- 1. Remove the antibody-coated microplate modules (marked "A") from the sealed foil pouch after the pouch has been equilibrated to room temperature. Affix the microplate modules to the supporting frame.
- 2. Wash the plare two times using $300 \ \mu$ L of wash buffer per well. After each wash, remove any remaining solution by inverting and tapping the plate firmly on a clean paper towel.
- 3. In each well, dispense 45 µL of sample diluent (marked "C").
- In each well, dispense 50 μL of guinea pig anti-leptin serum (marked "D").
- 5. Pipette 5 µL samples (or 0, 0.2, 0.4, 0.8, 1.6, 3.2, 6.4 and 12.8 ng/mL working rat leptin standards) into the wells.

6. Cover the microplate with the plastic microplate cover and incubate overnight (16 – 20 hours) at 4°C.

Second reaction:

- Aspirate well contents and wash five times using 300 µL of wash buffer per well. After each wash, remove any remaining solution by inverting and tapping the plate firmly on a clean paper towel.
- 8. Dispense 100 µL per well of anti-guinea pig IgG enzyme conjugate.
- 9. Cover the microplate with the plastic microplate cover and incubate for 3 hours at 4°C.

Third reaction:

- Aspirate well contents and wash seven times using 300 µL of wash buffer per well. After each wash, remove any remaining solution by inverting and tapping the plate firmly on a clean paper towel.
- Immediately dispense 100 µL per well of enzyme substrate solution (marked "G") and react for 30 minutes at room temperature. During the enzyme reaction, avoid exposing the microplate to light.

Note: Do not cover the microplate with aluminum foil.

- 12. Stop the enzyme reaction by adding 100 μL per well of enzyme reaction stop solution (marked "H").
- 13. Measure absorbance within 30 minutes using a plate reader. (Measure A_{450} values and subtract A_{630} values).


Note: Each standard and sample should be assayed in duplicate. It is also recommended that a 10 μ L or better precision pipette be used when dispensing small volumes (5 μ L).

I.4. Determining the leptin concentration

- 1. Determine the mean absorbance for each set of duplicate standards or samples.
 - **Note**: If individual absorbance values differ from the mean by greater than 20%, performing the assay again is recommended. The mean absorbance of the 0 ng/mL standard should be less than 0.1.
- 2. Using semi-log graph paper, construct the leptin standard curve by plotting the mean absorbance value for each standard on the Y axis versus the corresponding standard rat leptin concentration on the X axis. Figure 1 is an example of a typical standard curve generated by the assay.

Note: A standard curve should be plotted every time the assay is performed.

- 3. Rat leptin concentrations in the samples are interpolated using the standard curve and mean absorbance values for each sample.
 - **Note**: Samples with a high leptin concentration (12.8 ng/mL or higher) should be diluted with the sample diluent and rerun.



Rat Leptin (ng/ml)

Figure 1: Typical rat leptin standard curve

J.Appendix

J.1. Performance characteristics

- 1. Precision: The intra-assay precision: $C.V. \le 10\%$ The inter-assay precision: $C.V. \le 10\%$
- 2. Recovery: When rat leptin was spiked in rat serum sample, the recovery was $100\% \pm 20\%$.

J.2. Rat leptin recovery test

To determine the recovery of rat leptin, a minimum of three sample assays should be performed as illustrated in Table 3 (A, C, D or B, C, E).

				-		
Unknown sample (µL)	ample standard		Guinea pig anti-leptin serum (µL)	Total volume (µL)	Actua value (ng/m	Э
0	5 (1.6 ng/ml)	45	50	100	1.46	Α
0	5 (3.2 ng/ml)	45	50	100	3.09	В
5	0 (0 ng/ml)	45	50	100	2.80	С
5	5 (1.6 ng/ml)	40	50	100	3.53	D
5	5 (3.2 ng/ml)	40	50	100	5.19	Ε

TABLE 3 Example worksheet using a 5 µL sample

Once the assays have been performed, use either of the calculations below to determine the leptin recovery:

Calculation 1

Recovery (%) =
$$\frac{D (3.53 \text{ ng/mL})}{A (1.46 \text{ ng/mL})+C (2.80 \text{ ng/mL})} \times 100=82.9$$

Calculation 2

Recovery (%) =
$$\frac{\mathbf{E} (5.19 \text{ ng/mL})}{\mathbf{B} (3.09 \text{ ng/mL}) + \mathbf{C} (2.80 \text{ ng/mL})} \times 100 = \mathbf{88.1}$$

J.3. Summary of reagent preparation

Reagent	Preparation Procedure	
A: Antibody-coated Microplate	Ready to use	
B: Rat Leptin Standard,	Dilute with 100 µL of Sample	
Lyophilized	Diluent (marked " C ")	
C: Sample Diluent	Ready to use	
D: Guinea Pig Anti-Leptin Serum	Ready to use	
E: Anti-Guinea Pig IgG Enzyme	Mix the bottle of E with the	
Conjugate Stock Solution	bottle of F and mix	
F : Enzyme Conjugate Diluent	completely**	
G: Enzyme Substrate (TMB)	Roody to upo	
Solution	Ready to use	
H: Enzyme Reaction Stop	Ready to use	
Solution (1 N Sulfuric Acid)		
I: Wash Buffer Stock Solution	Bring contents of the bottle to	
(20X concentrate)	1 L with water*	

TABLE 4 – Summary of reagent preparation

Note: All reagents should be brought to room temperature (18-25°C) prior to use.

* Distilled or deionized water.

** Not needed till the second day of the assay.

J.4. Summary of Rat Leptin ELISA kit assay

Affix the Antibody-coated Microplate (marked "A") to the frame Wash each well two times with wash buffer* Dispense 45 µL of Sample Diluent (marked "C") per well Dispense 50 µL of Guinea Pig Anti-Leptin Serum (marked "D") per well Pipette 5 µL of the sample (or working rat leptin standard) per well Incubate the microplate overnight (16 -20 hours) at 4°C Wash each well five times with wash buffer* Dispense 100 µL of anti-guinea pig IgG enzyme conjugate per well Incubate the microplate for 3 hours at 4°C Wash each well seven times with wash buffer* Dispense 100 µL of Enzyme Substrate Solution (marked "G") per well Incubate microplate for 30 min at room temperature while avoiding exposure to light. Stop the enzyme reaction by adding 100 µL of Enzyme Reaction Stop Solution (marked "H") per well Measure A_{450} and subtract A_{630} values within 30 min Calculate leptin concentrations using the standard curve

* Each well should be washed with 300 μ L of wash buffer. Aspirate the wells completely so all excess solution is removed.

Warranty

Crystal Chem Inc. makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. Buyer assumes all risk and liability resulting from the use of this product.

THERE IS NO WARRANTY OF MERCHANTABILITY OF THE PRODUCTS, OR THAT SUCH PRODUCTS ARE FIT FOR ANY PARTICULAR PURPOSE. CRYSTAL CHEM INC'S LIABILITY SHALL NOT EXCEED THE RETURN OF THE PURCHASE PRICE, AND UNDER NO CIRCUMSTANCES SHALL CRYSTAL CHEM INC. BE LIABLE FOR SPECIAL OR CONSEQUENTIAL DAMAGES, OR EXPENSES ARISING DIRECTLY OR INDIRECTLY FROM THE USE OF THIS PRODUCT.